



Certificate of Analysis

1.06007.2500 Methanol gradient grade for liquid chromatography LiChrosolv®
Reag. Ph Eur
Batch I790007

Batch Values

Purity (GC)	≥ 99.9	%
Identity (IR)	conforms	
Evaporation residue	≤ 2.0	mg/l
Water	≤ 0.02	%
Colour	≤ 10	Hazen
Density (d 20 °C/20 °C)	0.793	
Boiling point	64 - 65	°C
Acidity	≤ 0.0002	meq/g
Alkalinity	≤ 0.0002	meq/g
Gradient grade (at 235 nm)	≤ 2.0	mAU
Gradient grade (at 254 nm)	≤ 1.0	mAU
Fluorescence (as quinine at 254 nm)	≤ 1.0	ppb
Fluorescence (as quinine at 365 nm)	≤ 0.5	ppb
Transmission (at 210 nm)	≥ 20	%
Transmission (at 220 nm)	≥ 60	%
Transmission (at 230 nm)	≥ 75	%
Transmission (at 235 nm)	≥ 83	%
Transmission (at 250 nm)	≥ 95	%
Transmission (from 260 nm)	≥ 98	%
Absorbance (at 225 nm)	0.131	

Filtered by 0.2 µm filter.

Suitable for UPLC / UHPLC /Ultra HPLC - instruments.

Conforms to Methanol R1 and R2 according to Reag. Ph Eur;
conforms to the requirements of ACS for liquid chromatography suitability.

Date of release (DD.MM.YYYY) 13.07.2015

Minimum shelf life (DD.MM.YYYY) 31.07.2018

Dr. Michael Savelsberg

Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.